Corporate Accreditation No 63

Accredited for compliance with ISO/IEC 17025 - Testing

Sydney WATER
Laboratory Services

Delivery Address: Sydney Water Corporation
51 Hermitage Rd
West Ryde NSW 2114

Telephone: (02) 9800 6935
Email: analyticalservices@sydneywater.com.au

Analytical Report 290781

Issue Date: 08/09/2023
Issued By : Sydney Water Laboratory Services

Attention: [redacted]
Customer: [redacted]
Customer ID: ZOEH_1

Address: [redacted]
Telephone: [redacted]
Email: [redacted]

CONTENTS

Sydney Water Approved Signatory

[redacted] Phycology Analyst
[redacted] Phycology Analyst
[redacted] Organics Analyst
[redacted] Organics Analyst

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

Sydney WATER Laboratory Services

SAMPLE SUMMARY

Client Sample ID	Sample Number	Sampling Procedure	Date Sampled	Date Received	Date Authorised	Description
237636	L23069425	1	22/08/2023	25/08/2023	06/09/2023	B1 (WATER SAMPLE)
237637	L23069426	1	22/08/2023	25/08/2023	06/09/2023	B2 (WATER SAMPLE)
237638	L23069427	1	22/08/2023	25/08/2023	06/09/2023	B3 (WATER SAMPLE)
237639	L23069428	1	22/08/2023	25/08/2023	06/09/2023	E1 (WATER SAMPLE)
237640	L23069429	1	22/08/2023	25/08/2023	06/09/2023	E2 (WATER SAMPLE)
237641	L23069430	1	22/08/2023	25/08/2023	06/09/2023	E3 (WATER SAMPLE)
237642	L23069431	1	22/08/2023	25/08/2023	06/09/2023	E4 (WATER SAMPLE)
237643	L23069432	1	22/08/2023	25/08/2023	06/09/2023	E5 (WATER SAMPLE)
237644	L23069433	1	22/08/2023	25/08/2023	01/09/2023	B1 (WATER SAMPLE)
237645	L23069434	1	22/08/2023	25/08/2023	01/09/2023	B2 (WATER SAMPLE)
237646	L23069435	1	22/08/2023	25/08/2023	01/09/2023	B3 (WATER SAMPLE)
237647	L23069436	1	22/08/2023	25/08/2023	05/09/2023	E1 (WATER SAMPLE)
237648	L23069437	1	22/08/2023	25/08/2023	05/09/2023	E2 (WATER SAMPLE)
237649	L23069438	1	22/08/2023	25/08/2023	01/09/2023	E3 (WATER SAMPLE)
237650	L23069439	1	22/08/2023	25/08/2023	06/09/2023	E4 (WATER SAMPLE)
237651	L23069440	1	22/08/2023	25/08/2023	07/09/2023	E5 (WATER SAMPLE)

Sampling procedures

1 Samples analysed as received.
2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
3 Samples collected as per FS procedures SAWI 070.
4 Results reported as received from WNSW.

Sydney WATER Laboratory Services

ANALYTICAL RESULTS

Client Sample ID		237636	237637	237638	237639	237640	237641	237642	237643
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069425	L23069426	L23069427	L23069428	L23069429	L23069430	L23069431	L23069432
ORGANICS									
TC0049DW : Algal Toxins									
Cylindrospermopsin (extra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Cylindrospermopsin (intra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Anatoxin-a(intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Microcystin RR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05

Sydney WATER Laboratory Services

Client Sample ID		237636	237637	237638	237639	237640	237641	237642	237643
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069425	L23069426	L23069427	L23069428	L23069429	L23069430	L23069431	L23069432
ORGANICS									
TC0049DW : Algal Toxins(Continued)									
Microcystin LR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Cylindrospermopsin(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Nodularin (total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Date of Performance	DD/MM/YY	28/08/23	28/08/23	28/08/23	28/08/23	28/08/23	28/08/23	28/08/23	28/08/23
TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS									

Sydney WATER Laboratory Services

Client Sample ID		237636	237637	237638	237639	237640	237641	237642	237643
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069425	L23069426	L23069427	L23069428	L23069429	L23069430	L23069431	L23069432
ORGANICS									
TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)									
Saxitoxin	ug/L	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4
Neosaxitoxin	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
dcSTX	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
C2	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX4	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX3	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX5	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcNeo	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX3	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX6	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX2	ug/L	<1	<1	<1	<1	<1	<1	<1	<1
GTX1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5

Sydney WATER Laboratory Services

Client Sample ID		237636	237637	237638	237639	237640	237641	237642	237643
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069425	L23069426	L23069427	L23069428	L23069429	L23069430	L23069431	L23069432
ORGANICS									
TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)									
C1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX2	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
Date of Performance	DD/MM/YY	04/09/23	04/09/23	04/09/23	04/09/23	04/09/23	04/09/23	04/09/23	04/09/23

Client Sample ID		237644	237645	237646	237647	237648	237649	237650	237651
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069433	L23069434	L23069435	L23069436	L23069437	L23069438	L23069439	L23069440
ALGAL									
MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes									
Blue Green ASU	ASU/mL	1369	1589	950.8	2079	1442	1635	925.6	1225
Blue Green Biovol	mm3/L	0.403	0.377	0.225	0.6	0.388	0.462	0.459	0.393
Potentially Toxic Blue Green	cells/mL	763	0	0	0	0	0	0	0

Sydney WATER Laboratory Services

Client Sample ID		237644	237645	237646	237647	237648	237649	237650	237651
Sampled Date		22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM	22/08/2023 12:00:00 AM
Sample Number		L23069433	L23069434	L23069435	L23069436	L23069437	L23069438	L23069439	L23069440
ALGAL									
MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes(Continued)									
Potentially Toxic Blue Green ASU	ASU/mL	51.1	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Potentially Toxic Blue Green Biovol	mm3/L	0.079	0.000	0.000	0.000	0.000	0.000	0.000	0.000
Total Blue Green	cells/mL	692300	836400	500400	1022000	740700	810400	460200	650600
MA91 : Individual Species Total Count, Total BioVol, Total ASU									
Algae Source*	N/A	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL
Date of Performance	DD/MM/YY	01/09/23 00:00	01/09/23 00:00	01/09/23 00:00	05/09/23 00:00	05/09/23 00:00	01/09/23 00:00	06/09/23 00:00	06/09/23 00:00

COMMENTS

Sample ID	Comment Level	Method	Test	Comment
L23069433	Method	MA91	-	Debris present in the sample.
L23069434	Method	MA91	-	Debris present in the sample.
L23069435	Method	MA91	-	Debris present in the sample.
L23069436	Method	MA91	-	Debris present in the sample.
L23069437	Method	MA91	-	Debris present in the sample.
L23069438	Method	MA91	-	Debris present in the sample.
L23069439	Method	MA91	-	Debris and cells resembling bacteria present in the sample.

* Indicates NATA accreditation does not cover the performance of this service

L23069440	**Method**	**MA91**	**-**	Debris present in the sample.

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LABORATORY QC RESULTS

N/A - Not Applicable
PQL - Practical Quantitation Limit
LOQ - Limit of Quantification
RPD - Relative Percent Difference
SPIKE/Positive Control - Addition of a known amount and concentration
Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1
Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1
Duplicate Precision = Not calculated - Result is outside test range

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LOQ	Blank	Control *Acceptance Criteria*	Spike *Acceptance Criteria*	Duplicate1	Duplicate2	RPD *Acceptance Criteria*
TC0049DW Anatoxin-a(extracellular)						
<0.1 ug/L	**<0.1**	**82** *50.0 - 120.0 ug/L*	**58 % Recovery** *50.0 - 130.0 % Recovery*	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0049DW Anatoxin-a(intracellular)						
<0.1 ug/L	**F**		**E**	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0049DW Anatoxin-a(total)						
<0.1 ug/L	**F**		**E**	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0049DW Cylindrospermopsin (extracellular)						
<0.05 ug/L	**<0.05**	**85** *50.0 - 120.0 ug/L*	**74 % Recovery** *50.0 - 130.0 % Recovery*	**0.13**	**0.13**	**B** *0.0 - 0.0 %*
TC0049DW Cylindrospermopsin (intracellular)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Cylindrospermopsin(total)						
<0.05 ug/L	**F**		**E**	**0.13**	**0.13**	**B** *0.0 - 0.0 %*

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LOQ	Blank	Control *Acceptance Criteria*	Spike *Acceptance Criteria*	Duplicate1	Duplicate2	RPD *Acceptance Criteria*
TC0049DW Microcystin LR(extracellular)						
<0.05 ug/L	**<0.05**	**72** *50.0 - 120.0 ug/L*	**51 % Recovery** *50.0 - 130.0 % Recovery*	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin LR(intracellular)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin LR(total)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin RR(extracellular)						
<0.05 ug/L	**<0.05**	**85** *50.0 - 120.0 ug/L*	**46 % Recovery** *50.0 - 130.0 % Recovery*	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin RR(intracellular)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin RR(total)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin YR(extracellular)						
<0.05 ug/L	**<0.05**	**85** *50.0 - 120.0 ug/L*	**55 % Recovery** *50.0 - 130.0 % Recovery*	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LOQ	Blank	Control *Acceptance Criteria*	Spike *Acceptance Criteria*	Duplicate1	Duplicate2	RPD *Acceptance Criteria*
TC0049DW Microcystin YR(intracellular)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Microcystin YR(total)						
<0.05 ug/L	**F**		**E**	**<0.05**	**<0.05**	**B** *0.0 - 0.0 %*
TC0049DW Nodularin (extracellular)						
<0.1 ug/L	**<0.1**	**81** *50.0 - 120.0 ug/L*	**59 % Recovery** *50.0 - 130.0 % Recovery*	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0049DW Nodularin (intracellular)						
<0.1 ug/L	**F**		**E**	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0049DW Nodularin (total)						
<0.1 ug/L	**F**		**E**	**<0.1**	**<0.1**	**B** *0.0 - 0.0 %*
TC0061DW C1						
<0.5 ug/L	**<0.5**	**73** *70.0 - 130.0 ug/L*	**100 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW C2						
<0.3 ug/L	**<0.3**	**110** *70.0 - 130.0 ug/L*	**130 % Recovery** *50.0 - 130.0 % Recovery*	**<0.3**	**<0.3**	**B** *0.0 - 0.0 %*

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LOQ	Blank	Control *Acceptance Criteria*	Spike *Acceptance Criteria*	Duplicate1	Duplicate2	RPD *Acceptance Criteria*
TC0061DW dcGTX2						
<0.5 ug/L	**<0.5**	**76** *70.0 - 130.0 ug/L*	**120 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW dcGTX3						
<0.3 ug/L	**<0.3**	**93** *70.0 - 130.0 ug/L*	**110 % Recovery** *50.0 - 130.0 % Recovery*	**<0.3**	**<0.3**	**B** *0.0 - 0.0 %*
TC0061DW dcNeo						
<0.5 ug/L	**<0.5**	**130** *70.0 - 130.0 ug/L*	**130 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW dcSTX						
<0.5 ug/L	**<0.5**	**72** *70.0 - 130.0 ug/L*	**100 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW GTX1						
<0.5 ug/L	**<0.5**	**82** *70.0 - 130.0 ug/L*	**130 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW GTX2						
<1 ug/L	**<1**	**76** *70.0 - 130.0 ug/L*	**120 % Recovery** *50.0 - 130.0 % Recovery*	**<1**	**<1**	**B** *0.0 - 0.0 %*
TC0061DW GTX3						
<0.5 ug/L	**<0.5**	**120** *70.0 - 130.0 ug/L*	**120 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

LOQ	Blank	Control *Acceptance Criteria*	Spike *Acceptance Criteria*	Duplicate1	Duplicate2	RPD *Acceptance Criteria*
TC0061DW GTX4						
<0.3 ug/L	**<0.3**	**110** *70.0 - 130.0 ug/L*	**120 % Recovery** *50.0 - 130.0 % Recovery*	**<0.3**	**<0.3**	**B** *0.0 - 0.0 %*
TC0061DW GTX5						
<0.5 ug/L	**<0.5**	**91** *70.0 - 130.0 ug/L*	**120 % Recovery** *50.0 - 130.0 % Recovery*	**<0.5**	**<0.5**	**B** *0.0 - 0.0 %*
TC0061DW GTX6						
<0.3 ug/L	**<0.3**	**110** *70.0 - 130.0 ug/L*	**130 % Recovery** *50.0 - 130.0 % Recovery*	**<0.3**	**<0.3**	**B** *0.0 - 0.0 %*
TC0061DW Neosaxitoxin						
<0.3 ug/L	**<0.3**	**100** *70.0 - 130.0 ug/L*	**130 % Recovery** *50.0 - 130.0 % Recovery*	**<0.3**	**<0.3**	**B** *0.0 - 0.0 %*
TC0061DW Saxitoxin						
<0.4 ug/L	**<0.4**	**82** *70.0 - 130.0 ug/L*	**110 % Recovery** *50.0 - 130.0 % Recovery*	**<0.4**	**<0.4**	**B** *0.0 - 0.0 %*

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

Extra Note:

F: Blank is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE Anatoxin-a(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

SPIKE Microcystin RR(extracellular) Spiking level less than 40 % of sample concentration

DUPLICATE Microcystin RR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcNeo B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcSTX B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX4 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX5 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX6 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Neosaxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

Sydney WATER Laboratory Services

DUPLICATE Saxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service