



Corporate Accreditation No 63
Accredited for compliance with ISO/IEC 17025 - Testing



Analytical Report 289922

Issue Date: 24/08/2023
Issued By : Sydney Water Laboratory Services

Delivery Address: Sydney Water Corporation
[Redacted]
[Redacted]
Telephone: [Redacted]
Email: [Redacted]

Attention: [Redacted]
Customer: Department of Planning and Environment
Customer ID: ZOEH_1

Address: [Redacted]
Telephone: [Redacted]
Email: [Redacted]

CONTENTS

- 1. Sydney Water Approved Signatory
- 2. Sample Summary
- 3. Analytical results
- 4. Comments
- 5. Laboratory QC results

Sydney Water Approved Signatory

[Redacted] Phycology Analyst	[Redacted] Phycology Analyst	[Redacted] Phycology Analyst
[Redacted] Organics Analyst	[Redacted] Organics Analyst	

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

SAMPLE SUMMARY

<u>Client Sample ID</u>	<u>Sample Number</u>	<u>Sampling Procedure</u>	<u>Date Sampled</u>	<u>Date Received</u>	<u>Date Authorised</u>	<u>Description</u>
236872	L23064335	1	26/07/2023	07/08/2023	11/08/2023	E1 (WATER SAMPLE)
236873	L23064336	1	26/07/2023	07/08/2023	11/08/2023	E2 (WATER SAMPLE)
236874	L23064337	1	26/07/2023	07/08/2023	11/08/2023	E3 (WATER SAMPLE)
236875	L23064338	1	26/07/2023	07/08/2023	11/08/2023	E4 (WATER SAMPLE)
236876	L23064339	1	26/07/2023	07/08/2023	11/08/2023	E5 (WATER SAMPLE)
236877	L23064340	1	26/07/2023	07/08/2023	11/08/2023	B1 (WATER SAMPLE)
236878	L23064341	1	26/07/2023	07/08/2023	11/08/2023	B2 (WATER SAMPLE)
236879	L23064342	1	26/07/2023	07/08/2023	11/08/2023	B3 (WATER SAMPLE)
236880	L23064343	1	26/07/2023	07/08/2023	23/08/2023	E1 (WATER SAMPLE)
236881	L23064344	1	26/07/2023	07/08/2023	23/08/2023	E2 (WATER SAMPLE)
236882	L23064345	1	26/07/2023	07/08/2023	23/08/2023	E3 (WATER SAMPLE)
236883	L23064346	1	26/07/2023	07/08/2023	20/08/2023	E4 (WATER SAMPLE)
236884	L23064347	1	26/07/2023	07/08/2023	18/08/2023	E5 (WATER SAMPLE)
236885	L23064348	1	26/07/2023	07/08/2023	23/08/2023	B1 (WATER SAMPLE)
236886	L23064349	1	26/07/2023	07/08/2023	18/08/2023	B2 (WATER SAMPLE)
236887	L23064350	1	26/07/2023	07/08/2023	18/08/2023	B3 (WATER SAMPLE)

Sampling procedures

- 1 Samples analysed as received.
- 2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
- 3 Samples collected as per FS procedures SAWI 070.
- 4 Results reported as received from WNSW.

ANALYTICAL RESULTS

Client Sample ID		236872	236873	236874	236875	236876	236877	236878	236879
Sampled Date		26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number		L23064335	L23064336	L23064337	L23064338	L23064339	L23064340	L23064341	L23064342
ORGANICS									
TC0049DW : Algal Toxins									
Cylindrospermopsin (extra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Cylindrospermopsin (intra cellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Anatoxin-a(intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (extracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Nodularin (intracellular)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Microcystin RR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(extracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05

Client Sample ID	236872	236873	236874	236875	236876	236877	236878	236879
Sampled Date	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number	L23064335	L23064336	L23064337	L23064338	L23064339	L23064340	L23064341	L23064342

ORGANICS

TC0049DW : Algal Toxins(Continued)

Microcystin LR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(intracellular)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Anatoxin-a(total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Cylindrospermopsin(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin LR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin RR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Microcystin YR(total)	ug/L	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05	<0.05
Nodularin (total)	ug/L	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1	<0.1
Date of Performance	DD/MM/YY	09/08/23	09/08/23	09/08/23	09/08/23	09/08/23	09/08/23	09/08/23	09/08/23

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS

Client Sample ID	236872	236873	236874	236875	236876	236877	236878	236879
Sampled Date	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number	L23064335	L23064336	L23064337	L23064338	L23064339	L23064340	L23064341	L23064342

ORGANICS

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)

Saxitoxin	ug/L	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4	<0.4
Neosaxitoxin	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
dcSTX	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
C2	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX4	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX3	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
GTX5	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcNeo	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX3	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX6	ug/L	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3	<0.3
GTX2	ug/L	<1	<1	<1	<1	<1	<1	<1	<1
GTX1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5

Client Sample ID	236872	236873	236874	236875	236876	236877	236878	236879
Sampled Date	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number	L23064335	L23064336	L23064337	L23064338	L23064339	L23064340	L23064341	L23064342

ORGANICS

TC0061DW : Paralytic Shellfish Toxins (PST) Analysis by UPLCMSMS(Continued)

C1	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
dcGTX2	ug/L	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5	<0.5
Date of Performance	DD/MM/YY	10/08/23	10/08/23	10/08/23	10/08/23	10/08/23	10/08/23	10/08/23	10/08/23

Client Sample ID	236880	236881	236882	236883	236884	236885	236886	236887
Sampled Date	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number	L23064343	L23064344	L23064345	L23064346	L23064347	L23064348	L23064349	L23064350

ALGAL

MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes

Blue Green ASU	ASU/mL	2474	2027	2059	523.9	1528	2533	2175	2204
Blue Green Biovol	mm3/L	0.639	0.535	1.07	0.531	0.401	0.719	0.516	0.523
Potentially Toxic Blue Green	cells/mL	0	0	208	208	0	0	0	0

Client Sample ID		236880	236881	236882	236883	236884	236885	236886	236887
Sampled Date		26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM	26/07/2023 12:00:00 AM
Sample Number		L23064343	L23064344	L23064345	L23064346	L23064347	L23064348	L23064349	L23064350
ALGAL									
MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes(Continued)									
Potentially Toxic Blue Green ASU	ASU/mL	0.00	0.00	13.9	13.9	0.00	0.00	0.00	0.00
Potentially Toxic Blue Green Biovol	mm3/L	0.000	0.000	0.021	0.021	0.000	0.000	0.000	0.000
Total Blue Green	cells/mL	1279000	1070000	911200	243900	793900	1281000	1145000	1160000
MA91 : Individual Species Total Count, Total BioVol, Total ASU									
Algae Source*	N/A	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL	EXTERNAL
Date of Performance	DD/MM/YY	22/08/23 00:00	22/08/23 00:00	22/08/23 00:00	18/08/23 00:00	18/08/23 00:00	22/08/23 00:00	18/08/23 00:00	18/08/23 00:00

COMMENTS

Sample ID	Comment Level	Method	Test	Comment
L23064343	Method	MA91	-	Debris present in the sample.
L23064344	Method	MA91	-	Debris present in the sample.
L23064345	Method	MA91	-	Debris present in the sample.
L23064346	Method	MA91	-	Debris present in the sample.
L23064347	Method	MA91	-	Debris present in the sample.
L23064348	Method	MA91	-	Debris present in the sample.
L23064349	Method	MA91	-	Debris present in the sample.

* Indicates NATA accreditation does not cover the performance of this service

L23064350	Method	MA91	-	Debris present in the sample.
-----------	--------	------	---	-------------------------------

* Indicates NATA accreditation does not cover the performance of this service

LABORATORY QC RESULTS

N/A - Not Applicable

PQL - Practical Quantitation Limit

LOQ - Limit of Quantification

RPD - Relative Percent Difference

SPIKE/Positive Control - Addition of a known amount and concentration

Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1

Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1

Duplicate Precision = Not calculated - Result is outside test range

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Anatoxin-a(extracellular)						
<0.1 ug/L	<0.1	78 50.0 - 120.0 ug/L	75 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Anatoxin-a(intracellular)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Anatoxin-a(total)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Cylindrospermopsin (extracellular)						
<0.05 ug/L	<0.05	94 50.0 - 120.0 ug/L	89 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Cylindrospermopsin (intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Cylindrospermopsin(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Microcystin LR(extracellular)						
<0.05 ug/L	<0.05	71 50.0 - 120.0 ug/L	61 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin LR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin LR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(extracellular)						
<0.05 ug/L	<0.05	75 50.0 - 120.0 ug/L	61 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin RR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin YR(extracellular)						
<0.05 ug/L	<0.05	79 50.0 - 120.0 ug/L	71 % Recovery 50.0 - 130.0 % Recovery	<0.05	<0.05	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0049DW Microcystin YR(intracellular)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Microcystin YR(total)						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
TC0049DW Nodularin (extracellular)						
<0.1 ug/L	<0.1	72 50.0 - 120.0 ug/L	64 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Nodularin (intracellular)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0049DW Nodularin (total)						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
TC0061DW C1						
<0.5 ug/L	<0.5	110 70.0 - 130.0 ug/L	71 % Recovery 50.0 - 130.0 % Recovery	<0.5	<0.5	B 0.0 - 0.0 %
TC0061DW C2						
<0.3 ug/L	<0.3	120 70.0 - 130.0 ug/L	76 % Recovery 50.0 - 130.0 % Recovery	<0.3	<0.3	B 0.0 - 0.0 %

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0061DW dcGTX2						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	75 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW dcGTX3						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	67 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW dcNeo						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	59 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW dcSTX						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	65 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX1						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	70 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX2						
<1 ug/L	<1	110 <i>70.0 - 130.0 ug/L</i>	77 % Recovery <i>50.0 - 130.0 % Recovery</i>	<1	<1	B <i>0.0 - 0.0 %</i>
TC0061DW GTX3						
<0.5 ug/L	<0.5	110 <i>70.0 - 130.0 ug/L</i>	88 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
TC0061DW GTX4						
<0.3 ug/L	<0.3	93 <i>70.0 - 130.0 ug/L</i>	65 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW GTX5						
<0.5 ug/L	<0.5	120 <i>70.0 - 130.0 ug/L</i>	53 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.5	<0.5	B <i>0.0 - 0.0 %</i>
TC0061DW GTX6						
<0.3 ug/L	<0.3	110 <i>70.0 - 130.0 ug/L</i>	94 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW Neosaxitoxin						
<0.3 ug/L	<0.3	100 <i>70.0 - 130.0 ug/L</i>	54 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.3	<0.3	B <i>0.0 - 0.0 %</i>
TC0061DW Saxitoxin						
<0.4 ug/L	<0.4	110 <i>70.0 - 130.0 ug/L</i>	62 % Recovery <i>50.0 - 130.0 % Recovery</i>	<0.4	<0.4	B <i>0.0 - 0.0 %</i>

* Indicates NATA accreditation does not cover the performance of this service

Extra Note:

F: Blank is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE Anatoxin-a(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE C2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcGTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcNeo B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE dcSTX B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX1 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX2 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX3 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX4 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX5 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE GTX6 B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Neosaxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Saxitoxin B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

* Indicates NATA accreditation does not cover the performance of this service

* Indicates NATA accreditation does not cover the performance of this service