



Corporate Accreditation No 63  
Accredited for compliance with ISO/IEC 17025 - Testing



## Analytical Report 284808

Issue Date: 19/05/2023  
Issued By : Sydney Water Laboratory Services

Delivery Address: Sydney Water Corporation  
51 Hermitage Rd  
West Ryde NSW 2114  
Telephone: (02) 9800 6935  
Email: [analyticalservices@sydneywater.com.au](mailto:analyticalservices@sydneywater.com.au)

Attention: [REDACTED]  
Customer: Department of Planning and Environment  
Customer ID: ZOEH\_1

Address: [REDACTED]  
Telephone: 02 9995 5031  
Email: [oeH.forensics@environment.nsw.gov.au](mailto:oeH.forensics@environment.nsw.gov.au)

### CONTENTS

- 1. Sydney Water Approved Signatory
- 2. Sample Summary
- 3. Analytical results
- 4. Comments
- 5. Laboratory QC results

### Sydney Water Approved Signatory

[REDACTED SIGNATURE AREA]

Where a result is required to meet a compliance limit or specification the associated uncertainty must be considered. Uncertainty estimates are available for all accredited test results.

**SAMPLE SUMMARY**

<u>Client Sample ID</u>	<u>Sample Number</u>	<u>Sampling Procedure</u>	<u>Date Sampled</u>	<u>Date Received</u>	<u>Date Authorised</u>	<u>Description</u>
232813	L23037183	1	30/04/2023	03/05/2023	18/05/2023	(WATER SAMPLE)
232815	L23037184	1	30/04/2023	03/05/2023	08/05/2023	(WATER SAMPLE)
232817	L23037185	1	30/04/2023	03/05/2023	18/05/2023	(WATER SAMPLE)
232819	L23037186	1	30/04/2023	03/05/2023	08/05/2023	(WATER SAMPLE)
232821	L23037187	1	30/04/2023	03/05/2023	18/05/2023	(WATER SAMPLE)
232823	L23037188	1	30/04/2023	03/05/2023	08/05/2023	(WATER SAMPLE)

**Sampling procedures**

- 1 Samples analysed as received.
- 2 Samples collected as per FS procedures SAWI 070, Excluding Oil & Grease which is collected as per clients instructions.
- 3 Samples collected as per FS procedures SAWI 070.
- 4 Results reported as received from WNSW.

**ANALYTICAL RESULTS**

<b>Client Sample ID</b>	232813	232815	232817	232819	232821	232823		
<b>Sampled Date</b>	30/04/2023 12:00:00 AM							
<b>Sample Number</b>	L23037183	L23037184	L23037185	L23037186	L23037187	L23037188		

**ALGAL**

**MA71CENT : Blue-Green ID & Enumeration, Including ASU & Biovolumes**

Blue Green ASU	ASU/mL	5261	-	5049	-	2176	-		
Blue Green Biovol	mm3/L	3.88	-	4.84	-	2.34	-		
Potentially Toxic Blue Green	cells/mL	9090	-	6810	-	3220	-		
Potentially Toxic Blue Green ASU	ASU/mL	454	-	348.8	-	121.9	-		
Potentially Toxic Blue Green Biovol	mm3/L	0.562	-	0.485	-	0.094	-		
Total Blue Green	cells/mL	2124000	-	1254000	-	616000	-		

**MA91 : Individual Species Total Count, Total BioVol, Total ASU**

Algae Source*	N/A	EXTERNAL	-	EXTERNAL	-	EXTERNAL	-		
Date of Performance	DD/MM/YY	18/05/23 00:00		18/05/23 00:00		18/05/23 00:00			

**ORGANICS**

**TC0049DW : Algal Toxins**

\* Indicates NATA accreditation does not cover the performance of this service

"-" = Not required or refer to Laboratory comment

<b>Client Sample ID</b>		232813	232815	232817	232819	232821	232823		
<b>Sampled Date</b>		30/04/2023 12:00:00 AM							
<b>Sample Number</b>		L23037183	L23037184	L23037185	L23037186	L23037187	L23037188		
<b>ORGANICS</b>									
TC0049DW : Algal Toxins(Continued)									
Anatoxin-a(extracellular)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Anatoxin-a(intracellular)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Anatoxin-a(total)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Cylindrospermopsin (extra cellular)	ug/L	-	0.82	-	0.66	-	0.48		
Cylindrospermopsin (intra cellular)	ug/L	-	0.05	-	0.14	-	0.10		
Cylindrospermopsin(total)	ug/L	-	0.87	-	0.80	-	0.58		
Microcystin LR(extracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin LR(intracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin LR(total)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin RR(extracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		

\* Indicates NATA accreditation does not cover the performance of this service

"-" = Not required or refer to Laboratory comment

<b>Client Sample ID</b>	232813	232815	232817	232819	232821	232823		
<b>Sampled Date</b>	30/04/2023 12:00:00 AM							
<b>Sample Number</b>	L23037183	L23037184	L23037185	L23037186	L23037187	L23037188		

**ORGANICS**

TC0049DW : Algal Toxins(Continued)

Microcystin RR(intracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin RR(total)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin YR(extracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin YR(intracellular)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Microcystin YR(total)	ug/L	-	<0.05	-	<0.05	-	<0.05		
Nodularin (extracellular)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Nodularin (intracellular)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Nodularin (total)	ug/L	-	<0.1	-	<0.1	-	<0.1		
Date of Performance	DD/MM/YY		05/05/23		05/05/23		05/05/23		

**COMMENTS**

<u>Sample ID</u>	<u>Comment Level</u>	<u>Method</u>	<u>Test</u>	<u>Comment</u>
L23037183	Method	MA91	-	Debris present in the sample.

\* Indicates NATA accreditation does not cover the performance of this service

"-" = Not required or refer to Laboratory comment

---

L23037185	Method	MA91	-	Debris present in the sample.
L23037187	Method	MA91	-	Debris present in the sample.

\* Indicates NATA accreditation does not cover the performance of this service

"-" = Not required or refer to Laboratory comment

**LABORATORY QC RESULTS**

N/A - Not Applicable

PQL - Practical Quantitation Limit

LOQ - Limit of Quantification

RPD - Relative Percent Difference

SPIKE/Positive Control - Addition of a known amount and concentration

Duplicate Precision = Accepted - Result 2 within 95% confidence limits of result 1

Duplicate Precision = Outlier - Result 2 outside 95% confidence limits of result 1

Duplicate Precision = Not calculated - Result is outside test range

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Anatoxin-a(extracellular)</b>						
<0.1 ug/L	<0.1	<b>91</b> 50.0 - 120.0 ug/L	<b>94 % Recovery</b> 50.0 - 130.0 % Recovery	<0.1	<0.1	<b>B</b> 0.0 - 0.0 %
<b>TC0049DW Anatoxin-a(intracellular)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	<b>B</b> 0.0 - 0.0 %
<b>TC0049DW Anatoxin-a(total)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	<b>B</b> 0.0 - 0.0 %
<b>TC0049DW Cyindrospermopsin (extracellular)</b>						
<0.05 ug/L	<0.05	<b>88</b> 50.0 - 120.0 ug/L	<b>91 % Recovery</b> 50.0 - 130.0 % Recovery	<0.05	<0.05	<b>B</b> 0.0 - 0.0 %
<b>TC0049DW Cyindrospermopsin (intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	<b>B</b> 0.0 - 0.0 %
<b>TC0049DW Cyindrospermopsin(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	<b>B</b> 0.0 - 0.0 %

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Microcystin LR(extracellular)</b>						
<0.05 ug/L	<0.05	<b>70</b> <i>50.0 - 120.0 ug/L</i>	<b>62 % Recovery</b> <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin LR(intracellular)</b>						
<0.05 ug/L	F		E	0.08	0.08	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin LR(total)</b>						
<0.05 ug/L	F		E	0.08	0.08	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin RR(extracellular)</b>						
<0.05 ug/L	<0.05	<b>71</b> <i>50.0 - 120.0 ug/L</i>	<b>70 % Recovery</b> <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin RR(intracellular)</b>						
<0.05 ug/L	F		E	0.08	0.08	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin RR(total)</b>						
<0.05 ug/L	F		E	0.08	0.08	<b>B</b> <i>0.0 - 0.0 %</i>
<b>TC0049DW Microcystin YR(extracellular)</b>						
<0.05 ug/L	<0.05	<b>74</b> <i>50.0 - 120.0 ug/L</i>	<b>66 % Recovery</b> <i>50.0 - 130.0 % Recovery</i>	<0.05	<0.05	<b>B</b> <i>0.0 - 0.0 %</i>

\* Indicates NATA accreditation does not cover the performance of this service

LOQ	Blank	Control <i>Acceptance Criteria</i>	Spike <i>Acceptance Criteria</i>	Duplicate1	Duplicate2	RPD <i>Acceptance Criteria</i>
<b>TC0049DW Microcystin YR(intracellular)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Microcystin YR(total)</b>						
<0.05 ug/L	F		E	<0.05	<0.05	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (extracellular)</b>						
<0.1 ug/L	<0.1	71 50.0 - 120.0 ug/L	70 % Recovery 50.0 - 130.0 % Recovery	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (intracellular)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %
<b>TC0049DW Nodularin (total)</b>						
<0.1 ug/L	F		E	<0.1	<0.1	B 0.0 - 0.0 %

\* Indicates NATA accreditation does not cover the performance of this service

## Extra Note:

F: Blank is not applicable for this analyte

E: Spike is not applicable for this analyte

DUPLICATE Anatoxin-a(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Anatoxin-a(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Cylindrospermopsin(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin LR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin RR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Microcystin YR(total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (extracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (intracellular) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

DUPLICATE Nodularin (total) B: Duplicate RPD reject criteria is not applicable, results are <10 times LOQ

\* Indicates NATA accreditation does not cover the performance of this service